



REGIONAL INSTITUTE OF MEDICAL SCIENCES, IMPHAL – 795 004
(An Autonomous Institute under the Ministry of Health & Family Welfare, Govt. of India)

CORRIGENDUM: Tender No: Equip/Tender/RIMS-12 dated 15th June 2012

No: Equip/Tender/RIMS-12

Imphal, the 20th June 2012

The following specifications shall replace the specifications contained in “ANNEXURE” of the tender No. Equip/Tender/RIMS-12 dt. 15th June 2012. Other terms & conditions and time line will remain the same.

Item No.:- (1) Ceiling suspended –Dual Head LED O.T. Light

Specifications:- (Approximate figures)

Maximum Illumination at 1m Working Distance	160.000 Lx
Dimensions of Light Head	700 mm x 700 mm
Colour Temperature	5000K(=Neutral Sun Light)
Colour Rendering Index	96
Focusable Light Field in 1m Working Distance	250mm-380mm
Illumination Depth	1530mm (830/700)
Dimming range	10-100%
Temperature increase near Surgeons Head & Operating Field	No Temperature increase. No IR-or UV Light
Unsterile Control Options-Touch Panel	On/Off/Satndby/Endo Light/Small to large Light Field Adjustment/Dimming
Sterile Control Options Via Touch Less Sensor Technology (preferable)	Large Light Field 50%/70%/100% Large Light Field 10%/30%/50% Small Light Field 50%/70%/100% Small Light Field 10%/30%/50% Endo Light-Green Standby/on/off
Lifetime of LEDs	>40.000h
Power Supply	220-240 V, or 110-120 V
Supply Voltage at Light Head	24V/DC
Protecting Class	Class 1
Protecting Class (supporting System)	IP 30
Protecting Class Light Housing	IP 42
Power Consumption per Light Housing	Max. 260 W
Frequency	50Hz-60 Hz

*Cost of the LED Bulb and other spare parts to be provided.

Item No. (2) :-

- | | |
|--------------------------------|---------------------|
| (a) Split type Air conditioner | :- 1.5 Ton Capacity |
| (b) Split type Air conditioner | :- 2 Ton Capacity |
| (c) Split type Air conditioner | :- 3 Ton Capacity |

Item No. (3):- Suitable Stabilizer for the above air –conditioners.

Prof. S. Sekharjit Singh
Director,
Regional Institute of Medical Sciences,
Imphal.