


Annexure- A

No. DEN/EQ/C/13

Imphal, the 30<sup>th</sup> May 2014.

List of the equipments for Dental College, RIMS, Imphal.

SI No	Name of items	Description
1	Manual Rotary Microtome (Imported model) with disposable blade facility	<ul style="list-style-type: none"><li>i. Manual rotary microtome (import model) with clockwise coarse feed, with specimen retraction &amp; trimming facility.</li><li>ii. Knife holder base and disposable blade adapter/ holder with lateral displacement features. including locking lever.</li><li>iii. Standard specimen clamp with adapter.</li><li>iv. Sectioning thickness selection 0.5 to 60 <math>\mu</math>m.</li><li>v. Sectioning thickness increments from 0.2 <math>\mu</math>m to 0.5 <math>\mu</math>m increments from 2 -10 <math>\mu</math>m in 1 <math>\mu</math>m increments form 10 -20 <math>\mu</math>m increments from 20 -6- <math>\mu</math>m increments.</li><li>vi. Vertical stroke length 60 <math>\mu</math>m.</li><li>vii. Horizontal specimen feed 27 mm</li><li>viii. Specimen orientation : 8 degree</li><li>ix. Trimming thickness selection : 10 <math>\mu</math>m &amp; 50 <math>\mu</math>m.</li><li>x. Specimen retraction : 200 <math>\mu</math>m.</li></ul>
2	Lead Screen	<ul style="list-style-type: none"><li>i. Height:- 6 feet</li><li>ii. Width:- 3 feet</li><li>iii. Viewing glass:- 8 inches x 10 inches</li><li>iv. Lead thickness:- min .25mm</li></ul>
3	Wax bath	
4	Knife sharpener	(Arkanas Stone)
5	Spencer knife	

  
(Prof. S. Sekharjit Singh)  
Director,  
Regional Institute of Medical Sciences,  
Imphal