

Annexure-A

Tender Notice No. *PROJ/E-RIMS-13*

Imphal, the 27th Feb 2014.

Due date 19th March 2014.

Specifications of Equipments:-

- 1) -20⁰C Deep Freezer
 - i) CFC free & Frost Free
 - ii) Vertical freezer with at least 265 litre capacity, with 2 or more shelves
 - iii) With digital temperature indicator and alarm
 - iv) Temperature adjustment should be $\pm 1^{\circ}\text{C}$
 - v) Outer body steel with powder coating or chrome plating
 - vi) Mounted on wheels
 - vii) Technical Manual should be provided along with the instrument
 - viii) Along with suitable UPS.

- 2) Rigid Nasal Endoscope with accessories.
 - i) 0 Degree:- a Hopkins II wide angel straight Forward telescope 0 degree, enlarged view, diameter 4mm, length 18 cm, autoclavable, filter optic light transmission incorporated, colour code: green.
 - ii) 30 Degree:- HOPKINS II wide Angle lateral Telescope 30 degree, enlarged view, diameter 2.7 mm, autoclavable. Fiber optic light transmission incorporated.


(Prof. S. Sekharjit Singh)

Director,
Regional Institute of Medical Sciences,
Imphal