

Corrigendum  
13/03/2020

**No. RIMS-61/Trauma/Major-E-2020 :** The scheduled dates and times of this office e-Tender Notice of even number dated 21<sup>st</sup> Feb 2020 for supply and installation of 800mA Digital Radiography machine and 32 Slice CT Scan machine for use of Trauma Care Centre of RIMS Hospital, Imphal have been amend and rescheduled as under :

- (1) The equipments should be BIS/USFDA/CE approved.
- (2) The firm registered under MSME / NSIC are exempted from EMD.
- (3) The last of tender was extend up to 4.30 p.m. of 20<sup>th</sup> March 2020.
- (4) Tender will be opened on 21<sup>st</sup> March 2020 at 4.35 p.m.
- (5) The amendments specifications are enclosed.
- (6) Other terms & conditions and annexure remain same and can be obtained from Institute website i.e. [www.rims.edu.in](http://www.rims.edu.in). and "<http://www.eprocure.gov.in>".



**Administrative Officer**  
**Regional Institute of Medical Sciences,**  
**Imphal.**

---

**Not to be published**

**No. RIMS-61/Trauma/Major-E-2020:**

**Dated 13<sup>th</sup> March 2020.**

- (1) Medical Superintendent RIMS Hospital, Imphal.
- (2) Prof. & HOD Surgery RIMS Imphal.
- (3) Prof. & HOD Radio diagnosis RIMS Imphal.
- (4) The System Administrator/Nodal Officer (e-tender) RIMS, Imphal for uploading the Corrigendum of Tender Notice to RIMS website.
- (5) Notice Board.

No.RIMS-61/Trauma/Major-E/2020

The following changes are made from the previous specifications and term and condition.

**Corrigendum for:-**

**Technical specification of 800mA or more Digital Radiography with Dual Detector.....**

1<sup>st</sup> para last line - The system should have two Detectors; one wireless / wired in the patient table and one fixed in the Bucky wall stand integrated with control console.

Sl. No. 1, iv

Image matrix size at least 2.3 X 2.8 K or more

Sl.No. 1, v

Analogue to Digital (A/D) conversion of 16 bits or better.

Sl.No.1, ix

Ingress protection should be IP X 6 or better.

Sl.No.2, ii

Output 80 KW or more (printing mistake was there in the previous specification)

Sl.No.3 ..... No change

Sl.No.3, iii .....No change

Sl.No.4, iii ..... No change

Sl.No.5, i

Height adjustable fixed four way floating table top with wired / wireless 35 X 43 cm or more sized FPD and Foot switch for controlling the table height, floating movement etc.

Sl.No.5, ii

It should have a carbon fiber top or equivalent.

Sl.No.6, iii ..... No change

Sl.No. 6, vi ..... No change

Sl.No. 7, ii

The Image Acquiring Software and Flat Panel Detector should be from the same manufacture for better synchronization and image output. Declaration and manufacturer details must be submitted with evidence.

Sl.No. 7, iv ..... No change

Sl.No.7, v

The Console PC must be of reputed brand and must have RAID 1 configuration to protect data. It should have 1 TB HDD and minimum 8 GB RAM.

Sl.No.9, i) a ..... No change


Sl.No.9, i) d ..... No change


Sl.No.9, i) g


System should be provided 2 computer system one for DICOM work station / viewing station. 50 packets of each 8 X 10, 10 X 12, and 11 X 14 film sizes should be provided.

**N.B.**

About the Extended Warranty of 5 years and change in GST rate :- Opinion of the Purchase Section is recommended.

  
13.3.2020


  
13.3.2020


  
13.3.20

No.RIMS-61/Trauma/Major-E/2020

Corrigendum for:-  
Technical Specification of 32 Slice CT Scanner.....

No gross changes is required.

  
13.3.2020

  
13.3.2020

  
13.3.20