



REGIONAL INSTITUTE OF MEDICAL SCIENCES: IMPHAL-795004
(An autonomous Institute under the Ministry of Health & Family Welfare, Government. of India.)

CORRIGENDUM OF TENDER

Imphal, the 28th December 2015

No. RIMS/PUR/EQUIP/TEN-12: The following technical specifications shall replace the specifications contained in annexure A.

- i. Wireless probes (both USG and Toco which are waterproff) and event marker { 2 (two) wireless proves for each machine is required}.
- ii. Hi-Tech colour TFT screen (minimum 10 inches) with tilt adjustment upto 90 degrees.
- iii. Tough Screen facility.
- iv. Light weight (mobile trolley with Antistatic surface & cupboard for storages is required)
- v. Automatic CTG reporting.
- vi. Interanal Battery backup of minimum 4 hours.
- vii. Patient data storage of 24 hours with printing facilities of FHR, TOCO and FM. The printing facility required in thermal printer with extra paper.

The Tender forms will be supplied upto 07-01-2016 and Sealed Tenders should be submitted by post/courier/hand upto 4.00 p.m of 11-01-2016. Sealed Tender will be opened on 15-01-2016 at 2.00 p.m and Other terms & conditions remain.


Sd/-
(Prof. S. Rita Devi)
Director,
Regional Institute of Medical Sciences,
Imphal.

Not to be published

Memo No. RIMS/PUR/EQUIP/TEN-12

Dated 28th Dec 2015

- (1) Medical Superintendent RIMS Hospital, Imphal.
- (2) Prof. & HOD Obst, & Gynae, RIMS Hospital, Imphal.
- (3) Deputy Director (Admn), RIMS, Imphal.
- (4) Chief Accounts Officer, RIMS, Imphal.
- (5) The System Administrator RIMS, Imphal for uploading the Corrigendum of Tender Notice to RIMS website.
- (6) Notice Board.


(Prof. S. Rita Devi)
Director,
Regional Institute of Medical Sciences,
Imphal.