



REGIONAL INSTITUTE OF MEDICAL SCIENCES: IMPHAL-795004
(An autonomous Institute under the Ministry of Health & Family Welfare, Government. of India.)

CORRIGENDUM OF TENDER

Imphal, the 11th August 2015

No. RIMS/E-13/42: The technical specifications for the Tender Notice of even number dated 23rd July 2015 for supply and installations of (a) Surgical Operating Microscope and (b) High Speed Electrical Drill System, for use of Neurosurgery RIMS Hospital, Imphal have been amend published through website.

(1) Other terms & conditions and annexure remain same and can be obtained from Institute website i.e. www.rims.edu.in.

(Prof. S. Rita Devi)
Director,
Regional Institute of Medical Sciences,
Imphal.

Corrigendum of Technical Specifications

No. RIMS/É-13/42

Imphal, the 11th Aug 2015

Item no.1:- Surgical Operating Microscope.

The following technical specifications are modified.

Tender Sl No.	Name of items	New Specifications
i.	Magnification	18 x Magnification or more
iv.	Field of view	14 mm to 140 mm with 12.5 X eye piece
v.	Focusing	Autofocus System. Motorized via multifocal lens activated through hand or foot switch or control panel. Also should have facility for manually adjustable override.
viii.	Light source	Xenon illumination with integrated Xenon back up
xii.	Binocular Tube	0-180 ^o

Item No.3:- High Speed Electrical Drill System

The following technical specifications are added.

Xv. Steering Duraguard should be at least 2 (two) or 3 (three) in nos.

xvi. Craniotomy attachment should be available in fixed and rotatable in three sizes.

xvii. To cater to Paediatrics' cases to adult with thick skull bone and should have a clear indicator on the open side of the tool at the time of drilling.



(Prof. S. Rita Devi)

Director,

Regional Institute of Medical Sciences,
Imphal