



REGIONAL INSTITUTE OF MEDICAL SCIENCES: IMPHAL-79500
(An Autonomous Institute under Ministry of Health & Family Welfare Govt. of India)

CORRIGENDUM

Imphal, the 21st July 2015

No. J/124/RIMS-2014 (pt): As per this office Tender Notice of even number dated 27th June 2015 for supply of Hospital Accessories for use of RIMS Hospital, Imphal, the error in the technical specifications of the following items mentioned below are changed as under.

Item name	Specifications
1. Suction machine	<ul style="list-style-type: none">• It should have minimum power of ¼ HP (0.25HP).• Type- Rotary Vane.• Lubrication- Oil (Continuous recirculating).• Bottle Type- Safe Autoclavable Poly Carbonate Jars.• Vacuum adjustment should be there.
2. Patient Trolley with Mattress	<ul style="list-style-type: none">• Overall approximate size should be (2030mm x 560mm x 820mm)(Length x Width x Height).• Rexine covered mattress with minimum thickness of 50 mm.
3. Wheel Chair	(Standard Fixed Type) Size 670mm(W) x 1120mm (D) x 920mm(L)
4. Cidex tray (Stainless Steel) a. 26''(L) x 6''(W) x 6''(H) b. 28'' (L) x 6''(W)x 6'' (H)	<ul style="list-style-type: none">• Stainless steel should be of 304 medical grade.
5. E.T Tube size 3, 3.5, 4, 4.5	<ul style="list-style-type: none">• It should be plain.
6. E.T Tube size 5, 5.5, 7, 8.5	<ul style="list-style-type: none">• It should be cuffed.
7. Rubberized Coir Mattress 76'' (L) x 36.3'' (W) x 4'' (thick)	Should be with i. Rexin Cover ii. 3'' of Coir & 1'' of Foam.
8. Protoscope	Small, Medium and Large

(Prof. S. Rita Devi)

Director,

Regional Institute of Medical Sciences,
Imphal

Copy to: -

- Medical Superintendent RIMS Hospital, Imphal
- System Administrator for uploading the above Corrigendum to the RIMS website.
- Notice Board