

REGIONAL INSTITUTE OF MEDICAL SCIENCES: IMPHAL-795004
(An autonomous Institute under the Ministry of Health & Family Welfare, Govt. of India.)

No. E/PMR/RIMS-13

Imphal the 3rd July 2015

Short Rate enquiry:

Subject:- Submission of Rate Quotations along with dealership/authorized Stockist / Manufacturer, certificate, catalogue, terms and conditions for supply and installation of an Electrosurgical unit.

Dear Sir,


I am directed to request you to kindly submit the rate quotation along with a copy of dealership/authorized stockist/manufacturer certificate, catalogue, terms and conditions for the following items on or before 13/07/15 and will be opened on 2.p.m of 14/07/15.

Electrosurgical unit - 1 no.

The following details, terms and conditions may please be fulfilled/noted

- (1) The price in words and figures should be indicated clearly.
- (2) The warranty period of the equipments is 3 a (three) year. If not mentioned, quotations shall be rejected.
- (3) The cover of the rate quotation should be superscribed as "rate quotation for supply of electrosurgical unit" (specifications at Annexure-A) for use of PMR RIMS Hospital, Imphal.
- (4) VAT charges for the items should be mentioned separately.
- (5) The rate quoted should be valid for one year from the due date of submitting the tenders.
- (6) Firm's TIN No., Registration No., PAN No., etc should be furnished. Pan card should be in the name of the firm or Proprietor's name.
- (7) The supply should be made within 4 weeks from the date of placing supply order.
- (8) Failure to supply should the items within the stipulated period and also within the validity period; the successful bidder/supplier is liable to be blacklisted.
- (9) Client list should be enclosed in the quotations.

Yours faithfully



(Y. Rajendra Singh)
Deputy Director (Admn),
Regional Institute of Medical Sciences,
Imphal.

Copy to:-

- (1) System Administrator for uploading the above short rate enquiry to the RIMS website.


Annexure-A

No. E/PMR/RIMS-13

Imphal the 3rd July 2015

Specifications for Electrosurgical Unit:-

- | | | |
|-------|--------------------------------------|------------------------|
| (i) | Frequency (output) | - 350 kHz |
| (ii) | Maximum cut output | - 200 watt at 500 ohms |
| (iii) | Maximum coagulation output | - upto 120 watts |
| (iv) | Safety regulation | - NESSY |
| (v) | System regulation | - 100-120/220-240 V |
| (vi) | Power input during maximum RF output | - 500 watts. |



(Y. Rajendra Singh)
Deputy Director (Admn)
Regional Institute of Medical Sciences,
Imphal.