



REGIONAL INSTITUTE OF MEDICAL SCIENCES: IMPHAL-795004
(An Autonomous Institute under Ministry of Health & Family Welfare, Govt. of India.)

Short Tender/Rate enquiry:

No. 065/VDL/MICRO/RIMS-2014

Imphal the 19th May 2018

Subject:- Submission of Rate Quotation along with dealership/authorized Stockiest /Manufacturer, certificate, terms and conditions for supply and installation of "Laminar Air Flow & - 20° C Deep Freezer"

Dear Sirs,

I am directed to request you to kindly submit the rate quotation along with a copy of dealership/authorized stokist/manufacturer certificate, terms & conditions for the supply and installation of 1 (one) no. each of Laminar Air Flow & - 20° C Deep Freezer (specification at annexure - A) for use of VRDL RIMS, Imphal on or before 26th May 2017 up to 4.00 p.m.

The following details, terms and Conditions may please be furnished.

- (1) The price in words and figures should be indicated clearly.
- (2) The warranty period of the battery is 1 (one) year.
- (3) The cover of the rate quotation should be subscribe "Rate quotation for supply of "Rate quotations for supply of "Laminar Air Flow & - 20° C Deep Freezer " for use of VRDL RIMS Imphal.
- (4) GST charges for the items should be mentioned clearly.
- (5) The rate tendered should be valid for one year from the due date of submitting the tenders.
- (6) Firm's GST No., Registration No., PAN No., etc.
- (7) The supplies should be made within 15 days from the date of placing supply order.
- (8) Failure to supply the items within the stipulated period and also within the validity period, the successful bidder/supplier is liable to be blacklisted.
- (9) Earnest money Rs.5,000/- (Rupees five thousand) only along-with rate quotation should be submitted in the form of Bank Draft from any Nationalized Bank/schedule bank in favour of Director RIMS, payable at Imphal. Interest on such Earnest Money will not be borne by us. The earnest money of unsuccessful firm will be refunded on demand. For the successful bidder, it will be retained as security deposit which shall be refundable after satisfactory completion of supplies. In the event of failure to supply the items with in the validity period and within the stipulated period from the date of placing supply order, the security Deposit shall be forfeited without any prior notice.

Yours faithfully


(E. Debendro Meitei)

Chief Accounts Officer-cum-Financial Advisor,
Regional Institute of Medical Sciences,
Imphal.

Copy to :-

(i)

System Administrator RIMS Imphal request for uploading the above notice to RIMS website.

Annexure – A

1. Laminar Air Flow

Specifications:

- a. Certification: EN12469 OR equivalent
- b. Design: Vertical laminar flow cabinet
- c. Approximately 4 Feet length X 2 feet depth, 304SS interior, epoxy/powder coated exterior Door fully closing with hinged sash.
- d. Support stand, solid one piece dished work surfaces
- e. Circulation: Down flow velocity of 40-70 fpm, ISO Class 5 or higher air
- f. Filters: HEPA filter, (99.99% efficient on particles 0.3 micron)/ULPA filter (efficiency of >99.999% at 0.1 to 0.3 micron sizes) provided with pre-filter)
- g. Light: UV and sufficient illumination for work space
- h. Gauges: For monitoring the condition of all HEPA filters as well as work space.
- i. Power Supply: Should include 210-240V/50-60 Hz

2. - 20^o C Vertical Deep Freezer

- Capacity : 600-700 litre
- Inner shelf: 6-9
- Refrigerant: CFC free Temperature control: Micro-processor controlled, Digital display with temperature resolution of 0.1°C.
- Alarms: Low/high temperature, power failure.
- Door closing and locking: Adjustment: self closing door with key door lock. Adjustable levelling feet standard (optional casters).
- Power supply: 210-240V/50-60 Hz


(E. Debendro Meitei)

Chief Accounts Officer-cum-Financial Advisor,
Regional Institute of Medical Sciences,
Imphal.