

REGIONAL INSTITUTE OF MEDICAL SCIENCES: IMPHAL-795004

(An autonomous Institute under the Ministry of Health & Family Welfare, Government. of India.)

No. RIMS-Pur/Nephro/E-19

Dated 12/11/2019


Corrigendum

All the equipments floated in the tender should be BIS/ISO/CE/FDA approved.

Technical Specifications of " Cardiac Monitor " :-

Point no.	Tender specifications	Amendment Specifications
12	Monitor should be able to store and recall at least 160 hours in both graphical & tabular format.	Monitor should be able to store and recall at least 120 hours in both graphical & tabular format.
18	Monitor should have facility for connecting high resolution large displays through latest HDMI ports.	Monitor should have facility for connecting high resolution large displays through latest HDMI/DVI-D/VGA Ports.
	The total nos. of Cardiac Monitor should be 6 (six).	

Other technical specifications and T & C remain same.


Head of Nephrology
RIMS Hospital, Imphal
Dr. T. Brojen Singh
Head ic & Assistant Professor
Department of Nephrology, RIMS, Imphal