

Annexure-A

No.87-RIMS-OPH-E-15

Imphal the 29th Dec 2015

(the equipment should be US FDA approved/CE Certified)

Item no.1:- Operating Microscope:

Specifications:-

- i. Achromatic optics with antireflex multicoating.
- ii. Motorized zoom system.
- iii. Good focusing range of atleast 50-60 mm.
- iv. Speed control for zoom and focus.
- v. 180 degree tiltable binocular tube f=170 mm.
- vi. Deep view depth of field management system.
- vii. Built in assistant microscope with independent magnification changer.
- viii. XY coupling with speed control.
- ix. SCI red reflex illumination.
- x. Xenon light source.
- xi. Magnetic clutches for all lock positions.
- xii. BIOM compatible.
- xiii. Beam splitter with still/video camera and monitor.



(Prof. S.Rita Devi)

Director

Regional Institute of Medical Sciences,
Imphal.