

Specification for Microtome

- Manual Rotary microtome
- Knife holder base
- Standard specimen clamp
- Disposable knife-80 mm long, 14 mm high

Automatic tissue processor 110-240 V/50-60 Hz:(2 Liter Capacity).

- One basket operation (tissue basket without vacuum/fume control).
- Unit with 12 processing stations, 10 reagent stations and 2 wax baths.
- Membrane key board and LCD screen.
- Alarms, Warning codes.
- Manual & automatic modes.
- Infiltration time: 99 hrs 59 min.
- Automatic tissue basket immersion

Water bath for paraffin section:

- Rectangular water bath
- High temperature consistency
- 200-240 V/50-60 Hz.

Dr. M. H. Meena
7/8/19

Dr. H. S. Singh
7/8/19

Professor & Head,
Department of Pathology
R.I.M.S., Imphal