

REGIONAL INSTITUTE OF MEDICAL SCIENCES: IMPHAL-795004

(An autonomous Institute under the Ministry of Health & Family Welfare, Government. of India.)

No. Pur-19/Plasma-Sterilizer/RIMS-18

dated 2/8/2019

Technical Specifications of " Plasma Sterilizer" as per Pre-bid meeting:-

Sl no.	Tender specifications	Amendment Specifications
7	It should have cassette type sterilant (min 4 to 7 cycles per cassette).	The Sterilant should be in a cassette /bottle/vial with hydrogen peroxide concentration.
15	It should have USB cycle data backup/ Ethernet connection. with 10" TFT touch LCD color screen.	It should have USB cycle data backup /Ethernet connection. with minimum 10" TFT touch LCD/LED color screen.
30	It should have SAL: 16-16 & international with safety standardization	It should have SAL: 16-16 & international with safety standardization/ other safety certificate given by govt.
32	The sterilizer should have approval/ validation claims from the leading medical reputed devices manufactures.	The sterilizer should have approval/ validation claims from the leading medical reputed devices manufactures/ Other relevant certification given by govt.
33	The System must be: US FDA & CE	The system should be US FDA/CE/BIS/ISO
34 (i)	Sterilant: Cassette	Sterilant: Cassette/bottle/vial

Other technical specifications and T & C remain same.

T.C. Members

- 1) En Singh M.S., RIMS.
- 2) Blues H.O.D, Surgery
3. N. Singh H.O.D, Medicine
4. Blues HOD Obs & Gynae
5. Blues HOD Orthopaedics.
6. Blues HOD Paediatrics.
7. Blues HOD Anaesthesia.