

## REGIONAL INSTITUTE OF MEDICAL SCIENCES: IMPHAL-795004 (An autonomous Institute under the Ministry of Health & Family Welfare, Government. of India.)

## **CORRIGENDUM**

## Imphal, the 22<sup>nd</sup> December 2018

No. Pur-20/Equip/RIMS-Hos-18: The technical specifications of this office Tender Notice of even number dated 20.11.2018 for supply & Installation of 6 sets of ICU Ventilator for use of RIMS, Imphal have been revised and the revised technical specifications enclosed at separate sheet.

- i. Due date, Time of submission of Tender is upto 4.00 p.m. of 7<sup>th</sup> January 2019 in lieu of 24<sup>th</sup> December 2018.
- ii. Date, Time of opening of Technical Bids is on 9<sup>th</sup> January 2019 in lieu of 26<sup>th</sup> December 2018.
- iii. Other terms & conditions remain same and can be obtained from Institute website i.e. <a href="www.rims.edu.in">www.rims.edu.in</a> & <a href="www.eprocure.gov.in">www.eprocure.gov.in</a>.

Note: - All the tender document received after extended date of 7th January 2019 will be presumed that bidders have rechecked the website and resubmit it as per corrigendum.

✓ (Kaikam Doungel)

Deputy Director (Admn),

Regional Institute of Medical Sciences,

Imphal.

## **ANNEXURE - A**

The following may be incorporated in the specification of ventilators for SICU.

- 1. Ventilator should have touch screen technology of 10 inches or more screen.
- 2. Tidal volume should be minimum of 10 ml i.e. range of 10 ml to 1000 ml or more.
- 3. Weight of ventilator may be omitted.
- 4. Ventilator should be turbine based or external compressor based (if external compressor, it should be external compressor of same manufacturer as the ventilator, with USFDA approved).
- 5. ETCO<sub>2</sub> upgradation facility.
- Breakdown should be connected immediately upto a maximum of 12 hours.
- 7. Demonstration of equipment before financial bid.
- 8. Availability of local service provider is preferred.

(Prof. K. lekhachandra)

(Prof S. Ravita).