

REGIONAL INSTITUTE OF MEDICAL SCIENCES: IMPHAL-795004
(An Autonomous Institute under Ministry of Health & Family Welfare, Govt. of India.)

No. E/PATH/96-RIMS

Imphal the 07/12/2015

Short Rate enquiry:

Subject:- Submission of Rate Quotations along with dealership/authorized Stockist /Manufacturer, certificate, terms and conditions for supply and installation of Centrifuge Machines.

Dear Sir,


I am directed to request you to kindly submit the rate quotation along with a copy of dealership /authorized stockist/manufacturer certificate, terms and conditions for the following items on or before ..18./12/..2015.. and bids will be opened on ..21/12/2015..... at 2.00 p.m.

Supply and installation of Centrifuge Machine- 3 (nos).
Specifications at Annexure-A

The following details, terms and Conditions may please be fulfilled/noted

- (1) The price in words and figures should be indicated clearly.
- (2) The warranty period of the equipments is 1 (one) year. If not mentioned, it will be assured that the warranty is for 1 year.
- (3) The cover of the rate quotation should be superscribed as "rate quotations for supply of Centrifuge Machine " for use of RIMS, Imphal".
- (4) VAT charges for the items should be mentioned separately.
- (5) The rate quoted should be valid for one year from the due date of submitting the tenders.
- (6) Firm's TIN No., Registration No., PAN No., etc should be furnished. Pan card should be in the name of firm or proprietor's name.
- (7) The supply should be made within 4 weeks from the date of placing supply order.
- (8) Failure to supply the items within the stipulated period and also within the validity period, the successful bidder/supplier is liable to be blacklisted.
- (9) Product catalogues should be enclosed along with the quotations.
- (10) Client list should be enclosed in the quotations.

Yours faithfully



(Y. Rajendra Singh)

Deputy Director (Admn),
Regional Institute of Medical Sciences,
Imphal.

Copy to:-

- (1) System Administrator for uploading the above Short rate enquiry to the RIMS website.

Annexure-A

No .E/PATH/96-RIMS

Imphal, the 07/12/2015.

Centrifuge Machines

3 nos.

Technical Specifications:-

Microprocessor controlled, swing out rotor.

- i. Maximum speed: 3,600 rpm.
- ii. The centrifuge must be capable of generating RCF in excess to 2,000 x g
- iii. Should be able to spin 8 tubes 13 x 75 mm or 13 x 100 mm at a time.
- iv. The system should have brush less induction drive.
- v. Dynamic braking for rapid stop without cell disturbance.
- vi. System should have safety features like lid lock and lid interlock.
- vii. System should have touch keypads for data entry and large LED displays for good visibility.
- viii. Noise level should be less than 60dB.
- ix. Electrical 220volts 50 Hz.

