

REGIONAL INSTITUTE OF MEDICAL SCIENCES: IMPHAL-795004
▸ (An Autonomous Institute under Ministry of Health & Family Welfare, Govt. of India.)

No. 67-Cobalt/M-R-14

Imphal the 15th Sept, 2015

Extension of Short Rate enquiry:

Subject:- Submission of Rate Quotations along with dealership/authorized Stockist /Manufacturer, certificates, terms and conditions for supply and installation of Radiation Survey Meter for Radiotherapy Department RIMS, Imphal.

Dear Sir,


I am directed to request you to kindly submit the rate quotation along with a copy of dealership /authorized stockist/manufacturer certificates, terms and conditions for the following items on or before 24/09/2015 and bids will be opened on 28/09/2015 at 2.00 p.m.

Radiation Survey Meter - 2 (two) nos.
(specifications at annexure-A)

The following details, terms and Conditions may please be fulfilled/noted

- (1) The price in words and figures should be indicated clearly.
- (2) The warranty period of the equipments is 3 (three) year. If not mentioned, tender shall be rejected.
- (3) The cover of the rate quotation should be superscribed as “**rate quotations for supply & installation of Radiation Survey Meter for use of RIMS, Imphal**”.
- (4) VAT charges for the items should be mentioned separately.
- (5) The rate quoted should be valid for one year from the due date of submitting the tenders.
- (6) Firm's TIN No., Registration No., PAN No., etc should be furnished. Pan card should be in the name of firm or proprietor's name.
- (7) The supply should be made within 4 weeks from the date of placing supply order.
- (8) Failure to supply the items within the stipulated period and also within the validity period, the successful bidder/supplier is liable to be blacklisted.
- (9) Product catalogues should be enclosed along with the quotations.
- (10) Client list should be enclosed in the quotations.
- (11) Firms are requested to quoted different models of different companies. For different brands/ make of the UPS, the firm should quote prices separately.

Yours faithfully



(Y. Rajendra Singh)

Deputy Director (Admn),
Regional Institute of Medical Sciences,
Imphal.

Copy to:-

✓ (1) System Administrator for uploading the above Extension of Short rate enquiry to the RIMS website.

Annexure- A


No.67-Cobalt/M-R-14

Imphal the 15th Sept, 2015

Name of the Item:- **Radiation Survey Meter**

Technical Specifications:-

- (i) Radiation detected – X, Beta and Gamma rays
- (ii) Range – 0.1 mR/hr to 200 mR/hr.
- (iii) Detector- Halogen quenched GM detector with shutter probe type
- (iv) Accuracy- $\pm 05\%$ in full scale.
- (v) Display- Analog bar scale with digital indication.
- (vi) Range selection- Automatic
- (vii) Battery indication – Facility for battery low indication
- (viii) Over range- Display
- (ix) Alarm system- Provision for alarm indication and rest
- (x) Measurement unit- CPM/R/hr/Sv/hr.



(Y.Rajendra Singh)
Deputy Director (Admn)
RIMS, Imphal.