

Annexure-A

No. RIMS/PUR/EQUIP/TEN-12

Imphal, the 7th December 2015

Item no.1:- Cardiotocograph (CTG) machine

10 (ten) nos.

(The machine should be US FDA approved/CE certified)

- i. Wireless probes (both USG and Toco which are waterproof) and event marker.
- ii. Hi-Tech colour TFT screen (minimum 10 inches) with tilt adjustment upto 90 degrees.
- iii. Tough Screen facility.
- iv. Light weight.
- v. Automatic CTG reporting.
- vi. Battery backup of minimum 4 hours.
- vii. Patient data storage of 24 hours with printing facilities of FHR, TOCO and FM.



(Prof. S.Rita Devi)
Director
Regional Institute of Medical Sciences,
Imphal.